

Warning : Do not modify the product. It may cause damages or broken.

Indication : To use for removals deposits on diseased teeth, or removals carious dentine.

Caution to use

- # Wash and sterilize before use.
- # Must be replaced for each patient.
- # Do not use for any other purpose
- # Do not give excessive stress during use, it may cause breakage or bending.
- # After use, remove any blood, body fluids, tissue fragments, etc. from the product promptly.
- # Do not use in patients having allergies.

Regarding Cleaning and Sterilization

- # Autoclaves must be used under the condition
115°C / 30 mins 121°C / 15 mins 132°C / 10 mins
- # Avoid using strong alkaline or acidic detergents or disinfectants as they may discolor or corrode the product.
- # Do not use metal scrubbers or cleansers (polishing powder) when cleaning the product as they will damage the surface.
- # Do not use the following materials as they may corrode metal.
Sodium hypochlorite, Povidone-iodine, Formalin, Gluconic acid, Chlorhexidine,
Highly acidic water, Benzalkonium chloride, Benzethonium chloride, Phenol, etc.

Serial Number : Marked on the package

Manufacturing Date : Marked on the package

Storage

- # Do not store with moisture or with deposits
- # This product may rust if left in a humid place for a long time, so store it in a clean place with low humidity.
- # After sterilization, be sure to store and manage them properly in a clean place to prevent recontaminations.

Product Owner

Sun Dental Co., Ltd.

4-8-9 Minami Senba, Chuo-ku, Osaka, 542-0081 Japan

Importer / Provider in Thailand

CHEERFUL LAB CO., LTD.

2991/71 Ladphrao, Klongchan, Bangkok, 10240 Thailand

TEL 02-261-1070 email info.cheerful.lab@gmail.com